



Michael R. Pence
Governor
Jerome M. Adams, MD, MPH
State Health Commissioner

DATE: December 2, 2015
TO: All Local Health Departments
Attn: Chief Food Inspection Officer
FROM: Laurie Kidwell, RRT Supervisor
Food Protection Program
SUBJECT: Insulet Corporation – RECALL [Medical Device]

AFFECTED PRODUCT: OmniPod (Pod) Insulin Management System

SUMMARY: Unclassified Recall; This Notification is due to a slight increase in the reported cases in which the Pod's needle mechanism failed to deploy or there was a delay in the deployment of the needle mechanism.

OmniPods from the affected lots listed below were distributed to customers in September 2015:

<u>Distribution Catalog Number</u>	<u>Description</u>	<u>Lot Number</u>
		L41880
		L41881
		L41892
		L41895
		L41897
		L41898
		L41899
United States	POD-ZXP420	OmniPod®, Insulin Management System
		L41900
		L41901
		L41902
		L41903
		L41904
		L41905
		L41906
		L41907

SUGGESTED ACTION: For consumer inquiry only. Consumers with questions may contact Insulet Customer Care via telephone at 1-855-407-3729 at any time.

Recall -- Firm Press Release

FDA posts press releases and other notices of recalls and market withdrawals from the firms involved as a service to consumers, the media, and other interested parties. FDA does not endorse either the product or the company.



2 North Meridian Street • Indianapolis, IN 46204
317.233.1325 tdd 317.233.5577
www.statehealth.in.gov

To promote and provide essential public health services.

Insulet Corporation Issues Field Safety Notification of OmniPod Insulin Management System

FDA posts press releases and other notices of recalls and market withdrawals from the firms involved as a service to consumers, the media, and other interested parties. FDA does not endorse either the product or the company.

For Immediate Release

December 1, 2015

Contact

Consumers

Insulet Customer Care
1-855-407-3729
Firm Press Release

On November 2, 2015, Insulet Corporation (Insulet or the Company) initiated a lot-specific voluntary Field Safety Notification (Notification) for 15 lots of the OmniPod (Pod) which were distributed in the U.S. and three lots which were distributed internationally.

Insulet has notified its distributors and customers by email, FedEx, and phone calls beginning on November 2, 2015, as well as all applicable regulatory agencies, including U.S. Food and Drug Administration (FDA). Consumers with questions may contact Insulet Customer Care via telephone at 1-855-407-3729 at any time.

This Notification is due to a slight increase in the reported cases in which the Pod's needle mechanism failed to deploy or there was a delay in the deployment of the needle mechanism. The reported incidence of this product issue in the affected lots is approximately 1%-2%. Once this issue was recognized, the Company corrected the manufacturing process and implemented additional inspection steps.

This Notification does not affect the OmniPod Personal Diabetes Manager (PDM).

In the event a needle mechanism fails to deploy, the needle will not be inserted and insulin delivery will not begin. The interruption of insulin delivery may cause elevated blood glucose (hyperglycemia), which, if left untreated, can result in diabetic ketoacidosis (DKA). If a patient has activated a Pod and experiences unexpected elevated blood glucose levels, a healthcare professional should be consulted.

The affected Pod lots have resulted in 66 Medical Device Reports, of which three required medical intervention. **No serious injuries or deaths have been reported in patients using OmniPod devices from the affected lots.**

Consumers who have Pods from the affected lots should ensure the needle mechanism has deployed properly, and may contact Insulet Customer Care via telephone at 1-855-407-3729 at any time.

OmniPods from the affected lots listed below were distributed to customers in September 2015:

Distribution	Catalog Number	Description	Lot Number
			L41880
			L41881
			L41892
			L41895
			L41897
United States	POD-ZXP420	OmniPod®, Insulin Management System	L41898
			L41899
			L41900
			L41901
			L41902
			L41903
			L41904

Distribution	Catalog Number	Description	Lot Number
			L41905 L41906 L41907
International	14810	OmniPod®, Insulin Management System	L41908 L41910 F41935

Adverse reactions or quality problems experienced with the use of this product may be reported to the FDA's MedWatch Adverse Event Reporting program either online, by regular mail or by fax.

- Complete and submit the report Online: www.fda.gov/medwatch/report.htm.
- Regular Mail or Fax: Download form www.fda.gov/MedWatch/getforms.htm or call 1-800-332-1088 to request a reporting form, then complete and return to the address on the pre-addressed form, or submit by fax to 1-800-332-0178.

###

